Bureau of Health Care Quality & Compliance

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER:		(X2) MULTIPLE CONSTRUCTION A. BUILDING B. WING		(X3) DATE SURVEY COMPLETED			
NVN4977PRI				B. WING		10/02/2009			
NAME OF PROVIDER OR SUPPLIER WARM SPRINGS CORRECTIONAL CENTER			3301 E 5TH	EET ADDRESS, CITY, STATE, ZIP CODE 1 E 5TH STREET RSON CITY, NV 89701					
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)			ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)		(X5) COMPLETE DATE		
S 0000	(EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)		on ised ised ine diet of s and of the ces n quacy nto ds of (c) nd . gation d as s,	S 000					
S 115	NAC 449.325 Infections and Communicable Diseases 1. A hospital shall: (a) Provide a sanitary environment to avoid sources and transmission of infections and communicable diseases This Regulation is not met as evidenced by: Surveyor: 21044		S 115						

If deficiencies are cited, an approved plan of correction must be returned within 10 days after receipt of this statement of deficiencies.

TITLE

(X6) DATE

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION (X1) PROVIDER/SUPPLIER/C IDENTIFICATION NUMBE			(X2) MULTIPLE CONSTRUCTION A. BUILDING B. WING		(X3) DATE SURVEY COMPLETED			
NVN4977PRI			10/02			/2009		
WARM SPRINGS CORRECTIONAL CENTER 3301 E 5TH			DRESS, CITY, STATE, ZIP CODE TH STREET CITY, NV 89701					
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FUL REGULATORY OR LSC IDENTIFYING INFORMATIO		FULL	ID PREFIX TAG	PROVIDER'S PLAN OF COR (EACH CORRECTIVE ACTION CROSS-REFERENCED TO THE A DEFICIENCY)	SHOULD BE	(X5) COMPLETE DATE	
S 115	Based on record revicorrectional center fa inmates received tube. Findings include: Inmate #8 - The inmate one-step TB skin test file did not contain a serie widence of a two-step. Inmate #9 - The inmate evidence of a two-step.	ew on 10/2/09, the liled to ensure 3 of 10 erculosis (TB) testing ate's file contained a t dated 7/1/09. The insecond-step TB skin test. The test of	imate rest. in	S 115				
0.120	Equipment 2. A hospital which prepares, sterilizes and stores its supplies and equipment directly shall develop systems and standards that are consistent with: (a) The standards for the control of infection established by the infection control officer of the hospital This Regulation is not met as evidenced by: Surveyor: 21044 Based on observation, record review and interview on 10/2/09, the correctional center was not following policy and procedures regarding the sterilization of instruments. Findings include: A 4/1/09 policy titled, "Sterilization of medical and dental instruments" was reviewed. The policy indicated that biological testing was to be done at a minimum of once a week to ensure the proper function of the sterilization equipment. The policy							

NAME OF PROVIDER OR SUPPLIER WARM SPRINGS CORRECTIONAL CENTER STREET ADDRESS, CITY, STATE, JIP CODE 3301 E STH STREET CARSON CITY, NV 89701 STATE CARSON CITY, NV 89701 STATE CARSON CITY, NV 89701 STATE CROSS REFERENCED TO THE APPROPRIATE CROSS REFERENCED TO TH	AND PLAN OF CORRECTION IDENTIFICATION NU		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER:		(X2) MULTIPLE CONSTRUCTION A. BUILDING		I \ /	(X3) DATE SURVEY COMPLETED	
MARK OF PROVIDER OR SUPPLIER WARM SPRINGS CORRECTIONAL CENTER (A) 10 E STIR STREET ADDRESS CITY, STATE, 2IP CODE 3301 E STIR STREET CARON CITY, NV 93701 (B) 10 (EACH DEPCICIONY MUST SER PROCESSED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION) S 126 Continued From page 2 Indicated that Attest ampoules were to be run weekly to determine the efficiency of the sterilizer and the test results were to be logged. The Attest log revealed that no weekly testing of ampoules was conducted from 1/16/08 to 5/24/19, from 5/24/09 to 8/7/09 and from 8/30/09 to 9/30/09. The policy also revealed the sterilizer was to receive monthly maintenance to prevent the unit from falling prematurely. The dental technician was interviewed about the maintenance schedule of the sterilizer and she reported the sterilizer had a "safety check" done in July of 2009 as evidenced by a inspection sticker, but that the unit had never been serviced monthly. S 128 NAC 449.327 Sterile Supplies and Medical Equipment 2. A hospital which prepares, sterilizes and stores its supplies and equipment directly shall develop systems and standards that are consistent with: (c) When applicable, the manufacturer's guidelines for the use and maintenance of the sequipment. This Regulation is not met as evidenced by: Surveyor: 21044 Based on observation and record review on 10/2/09, the correctional center was not following manufacturer's guidelines for the use and maintenance of the sterilizer equipment. Findings include:			NVN4977PRI		B. WING		10/02/2009		
PREFIX TAG REGULATORY OR LSC IDENTIFYING INFORMATION) S 126 Continued From page 2 indicated that Attest ampoules were to be run weekly to determine the efficiency of the sterilizer and the test results were to be logged. The Attest tog revealed that now evely testing of ampoules was conducted from 1/16/08 to 5/24/09, from 5/24/09 to 8/7/09 and from 8/30/09 to 9/30/09. The policy also revealed the sterilizer was to receive monthly maintenance to prevent the unit from failing prematurely. The dental technician was interviewed about the maintenance schedule of the sterilizer and she reported the sterilizer had a "safety check" done in July of 2009 as evidenced by a inspection sticker, but that the unit had never been serviced monthly. S 128 NAC 449.327 Sterile Supplies and Medical Equipment 2. A hospital which prepares, sterilizes and stores its supplies and equipment directly shall develop systems and standards that are consistent with: (c) When applicable, the manufacturer's guidelines for the use and maintenance of the equipment. This Regulation is not met as evidenced by: Surveyor: 21044 Based on observation and record review on 10/2/09, the correctional center was not following manufacturer's guidelines for the use and maintenance of the sterilizer equipment. Findings include:	NAME OF PROVIDER OR SUPPLIER			3301 E 5TH	STREET				
indicated that Attest ampoules were to be run weekly to determine the efficiency of the sterilizer and the test results were to be logged. The Attest log revealed that no weekly testing of ampoules was conducted from 1/16/08 to 5/24/09, from 5/24/09 to 8/7/09 and from 8/30/09 to 9/30/09. The policy also revealed the sterilizer was to receive monthly maintenance to prevent the unit from failing prematurely. The dental technician was interviewed about the maintenance schedule of the sterilizer and she reported the sterilizer had a "safety check" done in July of 2009 as evidenced by a inspection sticker, but that the unit had never been serviced monthly. S 128 NAC 449.327 Sterile Supplies and Medical Equipment 2. A hospital which prepares, sterilizes and stores its supplies and equipment directly shall develop systems and standards that are consistent with: (c) When applicable, the manufacturer's guidelines for the use and maintenance of the equipment. This Regulation is not met as evidenced by: Surveyor: 21044 Based on observation and record review on 10/2/09, the correctional center was not following manufacturer's guidelines for the use and maintenance of the sterilizer equipment. Findings include:	PREFIX	(EACH DEFICIENCY MUST BE PRECEDED BY FU			PREFIX	PREFIX (EACH CORRECTIVE ACTION SHOULD BE TAG CROSS-REFERENCED TO THE APPROPRIATE		COMPLETE	
the instruments (clamps, scissors, etc.) inside the sterile packages were in the closed position. Review of the autoclave operation manual		SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FUL REGULATORY OR LSC IDENTIFYING INFORMATION Continued From page 2 indicated that Attest ampoules were to be run weekly to determine the efficiency of the steri and the test results were to be logged. The Attest log revealed that no weekly testing of ampoules was conducted from 1/16/08 to 5/24/09, from 5/24/09 to 8/7/09 and from 8/30 to 9/30/09. The policy also revealed the steri was to receive monthly maintenance to preve the unit from failing prematurely. The dental technician was interviewed about the maintenance schedule of the sterilizer and sh reported the sterilizer had a "safety check" do in July of 2009 as evidenced by a inspection sticker, but that the unit had never been servi monthly. NAC 449.327 Sterile Supplies and Medical Equipment 2. A hospital which prepares, sterilizes and stores its supplies and equipment directly sha develop systems and standards that are consistent with: (c) When applicable, the manufacturer's guidelines for the use and maintenance of the equipment. This Regulation is not met as evidenced by: Surveyor: 21044 Based on observation and record review on 10/2/09, the correctional center was not follow manufacturer's guidelines for the use and maintenance of the sterilizer equipment. Findings include: Multiple instrument packs were observed. All the instruments (clamps, scissors, etc.) inside		rilizer 80/09 rilizer rent I I I I I I I I I I I I I I I I I I I					

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER:		(X2) MULTIPLE CONSTRUCTION A. BUILDING		(X3) DATE SURVEY COMPLETED		
	NVN4977PRI			B. WING		10/02/2009		
			STREET ADDI	RESS, CITY, STA	ATE, ZIP CODE	10.02		
WARM SPRINGS CORRECTIONAL CENTER			3301 E 5TH CARSON C	STREET ITY, NV 8970	1			
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)			ID PREFIX TAG	PROVIDER'S PLAN OF CORREC' (EACH CORRECTIVE ACTION SHOU CROSS-REFERENCED TO THE APPR DEFICIENCY)	ACTION SHOULD BE COMPLET TO THE APPROPRIATE DATE		
S 128	Continued From page	e 3		S 128				
	revealed instructions be in an open position	that all instruments sho n when sterilized.	ould					
S 129	NAC 449.327 Sterile Equipment	Supplies and Medical		S 129				
	3. If the supplies and equipment are sterilized on the premises of a hospital, the process of sterilization must be supervised by a person who has received specialized training in the operation of the process of sterilization, including training in methods of testing the process to verify the efficiency of the process of sterilization. This Regulation is not met as evidenced by: Surveyor: 21044 Based on record review and interview on 10/2/09, the correctional center did not ensure medical staff received training on sterilization of instruments and the use of the instrument autoclave.							
	During an interview with the dental technician, she reported that she sterilized the majority of the instruments, but during her absences, the nurses sterilized instruments. The dental technician had evidence of sterilization training in a file. When asked about sterilization training of the nurses, the Director of Nursing reported the nurses did not have evidence of any sterilization training in their files.							
S 181	S 181 NAC 449.3385 Dietary Personnel		S 181					
	The dietary service must be under the direction of a registered dietitian or other professional person who; (a) Is qualified in the field of institutional							

PRINTED: 10/19/2009 FORM APPROVED Bureau of Health Care Quality & Compliance STATEMENT OF DEFICIENCIES (X1) PROVIDER/SUPPLIER/CLIA (X3) DATE SURVEY (X2) MULTIPLE CONSTRUCTION AND PLAN OF CORRECTION COMPLETED IDENTIFICATION NUMBER: A. BUILDING B. WING NVN4977PRI 10/02/2009 STREET ADDRESS, CITY, STATE, ZIP CODE NAME OF PROVIDER OR SUPPLIER 3301 E 5TH STREET WARM SPRINGS CORRECTIONAL CENTER **CARSON CITY, NV 89701** SUMMARY STATEMENT OF DEFICIENCIES PROVIDER'S PLAN OF CORRECTION (X4) ID ID (X5) COMPLETE (EACH DEFICIENCY MUST BE PRECEDED BY FULL (EACH CORRECTIVE ACTION SHOULD BE PREFIX **PREFIX** DATE REGULATORY OR LSC IDENTIFYING INFORMATION) CROSS-REFERENCED TO THE APPROPRIATE TAG TAG DEFICIENCY) S 181 Continued From page 4 S 181 management, nutritional sciences or hotel restaurant management; (b) Has completed an academic program in culinary arts; or (c) Is certified as a dietary manager by the Dietary Managers Association and has additional work experience with medical and therapeutic diets. 3. The director of the dietary service may be employed on a full-time or part-time basis, or as a consultant. This Regulation is not met as evidenced by: Surveyor: 21044 Based on record review and interview on 10/2/09, the correctional facility did not ensure the culinary department was under the direction of a registered dietician. Findings include: There was no documentation that in-service training for food service personnel had been provided by the dietitian consultant. Staff stated that the dietitian did not conduct in-service trainings for culinary staff for safe and sanitary food handling practices because she had never been to the correctional center and only reviewed menus for nutritional adequacy. S 290 S 290 NAC 449.361 Nursing Services 1. A hospital shall have a well-organized plan that provides for 24-hour nursing services. The

nursing services must be furnished or supervised

This Regulation is not met as evidenced by:

by a registered nurse.

Surveyor: 21044

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/G IDENTIFICATION NUMB	ROVIDER/SUPPLIER/CLIA ENTIFICATION NUMBER:		PLE CONSTRUCTION	(X3) DATE SURVEY COMPLETED			
NVN4977PRI				B. WING		10/02/2009			
NAME OF PROVIDER OR SUPPLIER WARM SPRINGS CORRECTIONAL CENTER			3301 E 5TH	EET ADDRESS, CITY, STATE, ZIP CODE D1 E 5TH STREET RSON CITY, NV 89701					
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FU REGULATORY OR LSC IDENTIFYING INFORMATION			ID PREFIX TAG	PROVIDER'S PLAN OF COF (EACH CORRECTIVE ACTION CROSS-REFERENCED TO THE DEFICIENCY)	(X5) COMPLETE DATE			
S 290	Description of the staff schedule, it was discovered the infirmary was not staffed with nurses during the night shift from 11:00PM to 7:00AM. The Director of Nursing (DON) reported that if there was a medical emergency or if an inmate needed medication, guards would call the nearby correctional center for the night shift registered nurse.		n o orted in	S 290					
S 297	97 NAC 449.361 Nursing Service 8. The chief administrative nurse shall define the policies, procedures and standards relating to the provision of nursing services and shall ensure that the members of the nursing staff carry out those policies, procedures and standards. The policies, procedures and standards must be documented and accessible to each member of the nursing staff in written or electronic form. The chief administrative nurse must approve each element of the policies, procedures and standards before the element may be used or printo effect. This Regulation is not met as evidenced by: Surveyor: 21044 Based on record review and interview on 10/2/2 the correctional center failed to follow policy regarding annual competency testing for nursing staff performing laboratory tests. Findings include:		to the re out he re out he re of . The ch or put	S 297					

PRINTED: 10/19/2009 FORM APPROVED Bureau of Health Care Quality & Compliance STATEMENT OF DEFICIENCIES (X3) DATE SURVEY (X1) PROVIDER/SUPPLIER/CLIA (X2) MULTIPLE CONSTRUCTION AND PLAN OF CORRECTION COMPLETED IDENTIFICATION NUMBER: A. BUILDING B. WING NVN4977PRI 10/02/2009 STREET ADDRESS, CITY, STATE, ZIP CODE NAME OF PROVIDER OR SUPPLIER 3301 E 5TH STREET WARM SPRINGS CORRECTIONAL CENTER **CARSON CITY, NV 89701** SUMMARY STATEMENT OF DEFICIENCIES PROVIDER'S PLAN OF CORRECTION (X5) (X4) ID ID (EACH DEFICIENCY MUST BE PRECEDED BY FULL (EACH CORRECTIVE ACTION SHOULD BE COMPLETE PREFIX PREFIX DATE REGULATORY OR LSC IDENTIFYING INFORMATION) CROSS-REFERENCED TO THE APPROPRIATE TAG TAG DEFICIENCY) S 297 Continued From page 6 S 297 competency testing log was reviewed that documented the annual competency testing for the nurses that performed laboratory duties such as urine testing, collecting specimens, drawing blood and the use of the glucometer. The last entry in the log was dated 7/7/08. Staff reported they were supposed to perform competency testing per policy, but had stopped doing the testing.

If deficiencies are cited, an approved plan of correction must be returned within 10 days after receipt of this statement of deficiencies.